



No. 047

**The Hongkong Electric Co. Ltd. - Lamma Power Station Laboratory**

香港電燈有限公司 - 南丫發電廠實驗室

**ADDRESS** : G/F. & 1/F., New Control Building, Lamma Power Station, Lamma Island, Hong Kong  
**地址** : 香港南丫島南丫發電廠新行政大樓地下及二樓

**AUTHORISED REPRESENTATIVE** : Mr. CHIU Chun Ming, Jack, Senior Generation Chemist  
**委任代表** : 趙鎮明先生

**ENQUIRY** : Mr. CHIU Chun Ming, Jack, Senior Generation Chemist  
**查詢** : 趙鎮明先生

Tel. 電話 : 2982 6202  
Fax. 傳真 : 2982 1654  
E-mail 電子郵箱 : jchiu@hec.com.hk

**CLIENTELE** : In-house laboratory, normally not available for public testing  
**服務對象** : 內部實驗室，通常只提供內部服務

**ACCREDITED TEST CATEGORIES** : Chemical Testing 化學測試  
**認可測試類別** : Environmental Testing 環境測試

**SCOPE OF ACCREDITATION** : As shown on the following pages 詳見後頁  
**認可範圍**

\* \* \* \* \*

**The Hongkong Electric Co. Ltd. - Lamma Power Station Laboratory**

香港電燈有限公司 - 南丫發電廠實驗室

TEST CATEGORY 測試類別	ITEMS, MATERIALS OR PRODUCTS TESTED 測試項目、材料或產品	SPECIFIC TESTS OR PROPERTIES MEASURED 特定測試或量度的特性	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED 規範、標準方法或應用技術
CHEMICAL TESTING 化學測試	Pulverized Fuel Ash	Loss-on-ignition  Moisture content  Fineness	BS 3892:Part 1:1997 clause 12.2  BS 3892:Part 1:1997 clause .5  BS 3892:Part 1:1997 clause 6

**The Hongkong Electric Co. Ltd. - Lamma Power Station Laboratory**

香港電燈有限公司 - 南丫發電廠實驗室

TEST CATEGORY 測試類別	ITEMS, MATERIALS OR PRODUCTS TESTED 測試項目、材料或產品	SPECIFIC TESTS OR PROPERTIES MEASURED 特定測試或量度的特性	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED 規範、標準方法或應用技術
ENVIRONMENTAL TESTING 環境測試	Water and Wastewater	Physical Examination :-  - pH Value  - Total Suspended Solids dried at 103 - 105°C	APHA 20e 4500H <sup>+</sup> B (Glass electrode)  APHA 17e 2540D