



Sam Hing Scales Factory Limited – Testing and Calibration Laboratory
三興製磅廠有限公司 – 測試及校正實驗室

ADDRESS : Room 402, 4/F, Fuk Hong Industrial Building, 60 Tong Mi Road,
地址 Mongkok, Kowloon, Hong Kong
香港九龍旺角塘尾道 60 號福康工業大廈 4 字樓 402 室

ENQUIRY : Mr Alex LEUNG Yu-fung, Quality Manager
查詢 梁宇鋒先生

TELEPHONE : 2393 1144
電話

FAX : 2789 3030
傳真

E-MAIL : lab@samhing.com.hk
電郵

WEBSITE ADDRESS : www.samhing.com.hk
網址

CLIENTELE : Public
服務對象 公眾



Scope of Accreditation
Registration Number HOKLAS 148
Page 1 of 3
Issue Date: 3 December 2021
Ref: HOKLAS148-16

Sam Hing Scales Factory Limited – Testing and Calibration Laboratory

三興製磅廠有限公司 – 測試及校正實驗室

ADDRESS : Room 402, 4/F, Fuk Hong Industrial Building, 60 Tong Mi Road, Mongkok, Kowloon,
地址 Hong Kong
香港九龍旺角塘尾道 60 號福康工業大廈 4 字樓 402 室

ACCREDITED TEST : Calibration Services 校正服務

CATEGORY
認可測試類別

Sam Hing Scales Factory Limited – Testing and Calibration Laboratory

三興製磅廠有限公司 – 測試及校正實驗室

Room 402, 4/F, Fuk Hong Industrial Building, 60 Tong Mi Road, Mongkok, Kowloon, Hong Kong

香港九龍旺角塘尾道 60 號福康工業大廈 4 字樓 402 室

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	SPECIFIC TEST OR PROPERTY MEASURED[®] 特定測試或量度的特性 [®]	CALIBRATION AND MEASUREMENT CAPABILITY (CMC)* 校準和測量能力*
Mass and related quantities - Balance	<p>Calibration of mass using following OIML Class standard weights</p> <p>(1) E2 weights from 1 mg to 10 kg, (2) F1 weights from 1 mg to 10 kg, (3) F2 weights from 1 mg to 50 kg, (4) M1 weights from 0.5 g to 50 kg (5) M2 weights from 0.5 g to 100 kg (6) M2 weights from 500 kg to 1000 kg in accordance with in-house procedure STC1 over the following ranges:</p> <p>5 g to 200 g above 200 g to 1000 g above 1000 g to 6000 g above 6 kg to 60 kg above 60 kg to 600 kg above 600 kg to 1000 kg</p> <p>On-site calibration of mass using following OIML Class standard weights</p> <p>(1) E2 weights from 1 mg to 10 kg, (2) F1 weights from 1 mg to 10 kg, (3) F2 weights from 1 mg to 50 kg, (4) M1 weights from 0.5 g to 50 kg (5) M2 weights from 0.5 g to 100 kg (6) M2 weights from 500 kg to 1000 kg in accordance with in-house procedure STC1 over the following ranges:</p> <p>5 g to 200 g above 200 g to 1000 g above 1000 g to 6000 g above 6 kg to 60 kg above 60 kg to 600 kg above 600 kg to 6000 kg above 6000 kg to 60000 kg above 60000 kg to 100000 kg</p>	<p>0.02 mg 0.2 mg 2 mg 4 mg 0.07 g 7 g</p> <p>0.02 mg 0.2 mg 2 mg 4 mg 0.07 g 7 g 2 kg 40 kg</p>

[®] Unless otherwise specified, accredited activities are conducted at the laboratory.

* The calibration uncertainty of a device under test, which is usually reported at 95% confidence level, depends on both the CMC of the laboratory and the performance of the device during calibration.

