



Sartorius Hong Kong Limited – Calibration Laboratory

賽多利斯香港有限公司 – 校正實驗室

ADDRESS : Unit 1012, 10/F, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, Hong Kong
地址 香港九龍觀塘榮業街2號振萬廣場10樓1012室

ENQUIRY : Ms KANG Ling, Laboratory Manager
查詢 江玲女士

TELEPHONE : 2774 2678
電話

FAX : 2766 3526
傳真

E-MAIL : Jasmine.Kang@SARTORIUS.com
電郵

WEBSITE ADDRESS : <http://www.sartorius.com/>
網址

CLIENTELE : Public
服務對象 公眾



Scope of Accreditation
Registration No. HOKLAS 306
Page 1 of 3
Issue Date: 30 August 2024
Ref: HOKLAS306-3

Sartorius Hong Kong Limited – Calibration Laboratory

賽多利斯香港有限公司 – 校正實驗室

ADDRESS : Unit 1012, 10/F, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, Hong Kong
地址 : 香港九龍觀塘榮業街2號振萬廣場10樓1012室

ACCREDITED TEST : Calibration Services 校正服務
CATEGORY
認可測試類別

Sartorius Hong Kong Limited – Calibration Laboratory

賽多利斯香港有限公司 – 校正實驗室

Unit 1012, 10/F, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, Hong Kong

香港九龍觀塘榮業街2號振萬廣場10樓1012室

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	SPECIFIC TEST OR PROPERTY MEASURED[®] 特定測試或量度的特性 [®]	CALIBRATION AND MEASUREMENT CAPABILITY (CMC)* 校準和測量能力*
Mass and related quantities (cont'd) - Volume - Single channel piston-operated pipette (Type A only)	Calibration for delivered volume in accordance with ISO 8655-6: 2002 and verification for maximum permissible systematic error and random error as specified in ISO 8655-2: 2002 over the following ranges: 1 µL to 5 µL above 5 µL to 10 µL above 10 µL to 20 µL above 20 µL to 100 µL above 100 µL to 500 µL above 500 µL to 1 mL above 1 mL to 5 mL above 5 mL to 10 mL	0.07 µL 0.1 µL 0.7 µL 1 µL 5 µL 7 µL 29 µL 50 µL

[®] Unless otherwise specified, accredited activities are conducted at the laboratory.

* The calibration uncertainty of a device under test, which is usually reported at 95% confidence level, depends on both the CMC of the laboratory and the performance of the device during calibration.