

Annex III

List of activities for which accreditation has been terminated

(Notification Letter No. HOKLAS140-40; Effective date of termination: 9 December 2024)

Building Diagnostic Consultants Limited

香港工程質量檢測中心有限公司
12/F, Asia Trade Centre, 79 Lei Muk Road, Kwai Chung, NT
香港新界葵涌梨木道 79 號亞洲貿易中心 12 樓

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	SPECIFIC TEST OR PROPERTY MEASURED® 特定測試或量度的特性®	CALIBRATION AND MEASUREMENT CAPABILITY (CMC)* 校準和測量能力*
Concrete (diagnostic)	Carbonation test	Hong Kong Housing Authority Materials Testing Services (2012/2014) for Maintenance & Building Materials Specification Part D Cl. 4.3.1 (Method 1 & Method 2)

Annex II

List of activities for which accreditation has been terminated

(Notification Letter No. HOKLAS140-36; Effective date of termination: 1 December 2021)

Building Diagnostic Consultants Limited

香港工程質量檢測中心有限公司

12/F, Asia Trade Centre, 79 Lei Muk Road, Kwai Chung, NT

香港新界葵涌梨木道 79 號亞洲貿易中心 12 樓

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	SPECIFIC TEST OR PROPERTY MEASURED@ 特定測試或量度的特性@	CALIBRATION AND MEASUREMENT CAPABILITY (CMC)* 校準和測量能力*
Vibration measurements		
- Vibration monitoring device	<p>Calibration of the acceleration, velocity and displacement in accordance with in-house method DOC/CAL01 at frequencies of 20 Hz, 30 Hz, 40 Hz, 50 Hz and 60 Hz and dynamic acceleration range from above 10000 mm/s² to 90000 mm/s² over the following ranges:</p> <p>- Acceleration measurements above 10000 mm/s² to 90000 mm/s²</p> <p>- Velocity measurements 5mm/s to 179 mm/s</p> <p>- Displacement measurements 0.1 mm to 0.35 mm</p>	<p>13%</p> <p>13%</p> <p>13%</p>

[@] Unless otherwise specified, accredited activities are conducted at the laboratory.

* The calibration uncertainty of a device under test, which is usually reported at 95% confidence level, depends on both the CMC of the laboratory and the performance of the device during calibration.