

## Annex II

### List of activities for which accreditation has been terminated

(Notification Letter No. HOKLAS140-36; Effective date of termination: 1 December 2021)

#### **Building Diagnostic Consultants Limited**

香港工程質量檢測中心有限公司

12/F, Asia Trade Centre, 79 Lei Muk Road, Kwai Chung, NT

香港新界葵涌梨木道 79 號亞洲貿易中心 12 樓

#### **Calibration Services 校正服務**

| ITEM TESTED OR MEASURED<br>測試或量度項目                          | SPECIFIC TEST OR<br>PROPERTY MEASURED@<br>特定測試或量度的特性@   | CALIBRATION AND MEASUREMENT<br>CAPABILITY (CMC)*<br>校準和測量能力*                      |
|---|---|---|
| Vibration measurements<br><br>- Vibration monitoring device | Calibration of the acceleration, velocity and displacement in accordance with in-house method DOC/CAL01 at frequencies of 20 Hz, 30 Hz, 40 Hz, 50 Hz and 60 Hz and dynamic acceleration range from above 10000 mm/s <sup>2</sup> to 90000 mm/s <sup>2</sup> over the following ranges:<br><br>- Acceleration measurements<br>above 10000 mm/s <sup>2</sup> to 90000 mm/s <sup>2</sup><br><br>- Velocity measurements<br>5mm/s to 179 mm/s<br><br>- Displacement measurements<br>0.1 mm to 0.35 mm | <br><br><br><br><br><br><br><br><br><br>13%<br><br><br><br>13%<br><br><br><br>13% |

@ Unless otherwise specified, accredited activities are conducted at the laboratory.

\* The calibration uncertainty of a device under test, which is usually reported at 95% confidence level, depends on both the CMC of the laboratory and the performance of the device during calibration.