Annex III

List of calibrations for which accreditation has been terminated

(Notification letter No.: HOKLAS299-6; Effective date of termination: 29 January 2026)

Gilson China Limited – Gilson Hong Kong Laboratory

Unit Nos. 1 & 2, 20/F., Remington Center, 23 Hung To Road, Kwun Tong, Kowloon, Hong Kong
香港九龍觀塘鴻圖道 23 號利登中心 20 樓 2001-02 室

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	**SPECIFIC TEST OR PROPERTY MEASURED@** 特定測試或量度的特性@	**CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*** 校準和測量能力*
Accreditation of the following calibrations has been terminated		
Mass and related quantities		
- Volume		
- Single channel piston-operated pipette (Type A only)	Calibration for delivered volume in accordance with ISO 8655-6: 2002 and verification for maximum permissible systematic error and random error as specified in ISO 8655-2: 2002	
	over the following ranges:	
	1 µL	0.09 µL
	above 1 µl to 10 µL	0.12 µL
	above 10 µL to 50 µL	0.5 µL
	above 50 µL to 100 µL	1.5 µL
	above 100 µL to 200 µL	4.0 µL
	above 200 µL to 1000 µL	7.4 µL

@ Unless otherwise specified, accredited activities are conducted at the laboratory.

* The calibration uncertainty of a device under test, which is usually reported at 95% confidence level, depends on both the CMC of the laboratory and the performance of the device during calibration.