Annex III

List of tests for which accreditation has been terminated

(Notification Letter No. HOKLAS302-3; Effective date of termination: 8 September 2022)

ESG Matters Limited - Acoustic Calibration Centre

益思智綠色金融科技有限公司-聲學校正實驗所

Unit 1818-19, 18/F, Tower A, Regent Centre, 63 Wo Yi Hop Road, Kwai Chung, Hong Kong

香港葵涌和宜合道 63 號麗晶中心 A 座 18 樓 1818-19 室

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	SPECIFIC TEST OR PROPERTY MEASURED [@] 特定測試或量度的特性 [@]	CALIBRATION AND MEASUREMENT CAPABILITY (CMC)* 校準和測量能力*
	The accreditation of the following calibration has be	en terminated
Acoustic measurement		
- Sound calibrator	Calibration of the following parameters at 1 kHz using comparison method in accordance with in-house test procedure ESG-Noise-003:	
	- Sound pressure level 94 dB 114 dB	0.13 dB 0.13 dB
	- Frequency accuracy	0.2 Hz
	- Total distortion at 94 dB	0.2 %
	- Total distortion at 114 dB	0.2 %

[®] Unless otherwise specified, accredited activities are conducted at the laboratory.

^{*} The calibration uncertainty of a device under test, which is usually reported at 95% confidence level, depends on both the CMC of the laboratory and the performance of the device during calibration.