

Annex II

List of calibration activities for which accreditation has been terminated (Notification Letter No. HOKLAS306-5; Effective date of termination: 31 March 2026)

Sartorius Hong Kong Limited – Calibration Laboratory

賽多利斯香港有限公司 – 校正實驗室

Unit 1012, 10/F, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, Hong Kong

香港九龍觀塘榮業街2號振萬廣場10樓1012室

Calibration Services 校正服務		
ITEM TESTED OR MEASURED 測試或量度項目	SPECIFIC TEST OR PROPERTY MEASURED® 特定測試或量度的特性®	CALIBRATION AND MEASUREMENT CAPABILITY (CMC)* 校準和測量能力*
The accreditation of the following calibration has been terminated		
Mass and related quantities (cont'd) - Volume - Single channel piston-operated pipette (Type A only)	Calibration for delivered volume in accordance with ISO 8655-6: 2002 and verification for maximum permissible systematic error and random error as specified in ISO 8655-2: 2002 over the following ranges: 1 µL to 5 µL above 5 µl to 10 µL above 10 µL to 20 µL above 20 µL to 100 µL above 100 µL to 500 µL above 500 µL to 1 mL above 1 mL to 5 mL above 5 mL to 10 mL	0.07 µL 0.1 µL 0.7 µL 1 µL 5 µL 7 µL 29 µL 50 µL